



Certificate of Analysis

PMM-Swab

Article No.: 885.0120

Lot No.: 008821

Production date 2024-Dec-02

Prelim. exp. date: 2025-Jul-30

Storage: 2°-25°C

This medium can be adjusted and / or supplemented according to the performance criteria required.

Test	Specifications	Results
Appearance	Clear, yellowish	conform
pH value	7,1 - 7,5	7,2
Irradiation dose	25,0 - 50,0 kGy	conform

Microorganisms (10 - 100 CFU)	Conditions	Specifications	Results
Staphylococcus aureus ATCC 6538	≤3 d, 30-35 °C	Good growth	conform
Escherichia coli ATCC 8739	≤3 d, 30-35 °C	Good growth	conform
Pseudomonas paraeruginosa ATCC 9027	≤3 d, 30-35 °C	Good growth	conform
Bacillus spizizenii ATCC 6633	≤3 d, 30-35 °C	Good growth	conform
Staphylococcus aureus ATCC 6538 + 20 µl Aerodesin 2000	20 - 24h, 30 - 35 °C	Good growth	conform
Candida albicans ATCC 10231	≤5 d, 20-25 °C	Good growth	conform
Aspergillus brasiliensis ATCC 16404	≤5 d, 20-25 °C	Good growth	conform

Sterility control: no growth (30-35°C, ≥7d)

Date: 2025-Feb-03

Signed: Dr. Anja Rüger
Head of QC

This document has been established electronically and is valid without signature.